



Ambi-ance

Ambidextrous Nitrile Gloves Powder-Free

Maximum comfort and performance with textured fingertips for improved dexterity, enhanced grip and tactile sensitivity. Soft stretch Nitrile for a comfortable custom fit. Puncture and tear resistant barrier. Beaded cuff for extra protection.

REF#		
19010	EXTRA SMALL	BLUE
REF#		
19011	SMALL	BLUE
REF#		
19012	MEDIUM	BLUE
REF#		
19013	LARGE	BLUE
REF#		
19014	EXTRA LARGE	BLUE

200/BOX

Product Name: Ambidextrous Nitrile Gloves Powder-Free

Latex: No

CE: Yes

Made in USA: No

Worth Knowing:

CLEARED FOR USE WITH CHEMOTHERAPY DRUGS. TEST CHEMOTHERAPY DRUG AND CONCENTRATION MINIMUM BREAKTHROUGH DETECTION TIME (Specimen 1/2/3) (Minutes)

Carboplatin, 10 mg/ml (10,000ppm) No breakthrough up to 240 min. Carmustine (BCNU), 3.3 mg/ml (3,300 ppm) 1.82 (3.96, 1.82, 5.98) Cisplatin, 1.0 mg/ml (1,000 ppm) No breakthrough up to 240 min. Cyclophosphamide (Cytoxan), 20.0 mg/ml (20,000 ppm) No breakthrough up to 240 min. Dacarbazine (DTIC), 10.0 mg/ml (10,000 ppm) No breakthrough up to 240 min. Doxorubicin Hydrochloride, 2.0 mg/ml (2,000 ppm) No breakthrough up to 240 min. Etoposide (Toposar) 20.0 mg/ml (20,000 ppm) No breakthrough up to 240 min. Fluorouracil 50.0 mg/ml (50,000 ppm) No breakthrough up to 240 min. Ifosfamide, 50.0 mg/ml (50,000 ppm) No breakthrough up to 240 min. Methotrexate, 25 mg/ml (25,000 ppm) No breakthrough up to 240 min Mitomycin C, 0.5 mg/ml (500 ppm) No breakthrough up to 240 min. Mitoxantrone, 2 mg/ml (2,000 ppm) No breakthrough up to 240 min. Paclitaxel (Taxol), 6.0 mg/ml (6,000 ppm) No breakthrough up to 240 min. Thiotepa, 10.0 mg/ml (10,000 ppm) 0.93 (3.09, 5.38, 0.93) Vincristine Sulfate, 1.0 mg/ml (1,000 ppm) No breakthrough up to 240 min.

Ambi-ance™ Ambidextrous Nitrile Gloves were tested with the above chemotherapy drug concentrations in accordance with ASTM D 6978-05, Standard practice assessment of Resistance of Medical Gloves to Permeation by Chemotherapy drugs. Note: Testing conditions used were intended to approximate the worse case conditions for clinical use. It is the user's responsibility to determine the applicability of these gloves for the intended use with chemotherapy drugs.

